COMMAND
MEDICAL PRODUCTS

ADVANCED WELDING AND THE EXPANDING USE OF POLYURETHANE IN MEDICAL DEVICE MANUFACTURING

Practical Considerations for Material Selection, Process Development, and Manufacturing Success

By David Bowling
Senior RF Manufacturing Engineer



EXECUTIVE SUMMARY

Polyurethane (PU) has become one of the fastest-growing materials in medical device manufacturing due to its exceptional combination of bio-compatibility, flexibility, durability, and design versatility. As medical device manufacturers continue to move away from traditional materials such as PVC and latex, polyurethane is increasingly being specified for applications ranging from fluid management systems and drug delivery devices to wound care products and wearable technologies.

The success of polyurethane-based devices depends not only on material selection but also on the ability to reliably join components without compromising performance. Welding technologies, particularly radio frequency (RF) welding, have emerged as preferred manufacturing methods due to their ability to create strong, repeatable, and hermetic bonds without the use of adhesives or solvents.

RF welding is especially well-suited for polyurethane because it generates heat internally within the material, allowing precise control of the weld zone while minimizing surface degradation. When properly designed and validated, RF welded polyurethane assemblies offer exceptional strength, leak resistance, and long-term reliability.

This paper examines why polyurethane continues to gain market share in medical devices, explores the factors that influence weld quality, and outlines practical recommendations for achieving robust and repeatable RF welding processes.

INTRODUCTION

Medical device manufacturers face increasing demands for performance, safety, bio-compatibility, and sustainability. These requirements are driving a shift toward advanced polymer materials capable of delivering improved functionality while supporting efficient manufacturing processes.

Among these materials, polyurethane has emerged as a preferred solution for a wide range of medical applications. Its ability to combine flexibility, toughness, clarity, chemical resistance, and bio-compatibility makes it attractive for both disposable and long-term use devices. As a result, polyurethane is increasingly replacing traditional materials such as PVC, latex, and certain silicone formulations in many applications.

At the same time, the growing use of polyurethane has created new manufacturing challenges. Medical devices often require complex assemblies consisting of films, tubing, reservoirs, bladders, balloons, and molded components that must be joined reliably and repeatedly. Selecting the appropriate joining technology is therefore critical to device performance and manufacturability.

Among available joining methods, RF welding continues to demonstrate significant advantages for polyurethane materials. Understanding the relationship between polyurethane properties and RF welding performance allows device designers and manufacturers to optimize product reliability while reducing manufacturing complexity.




Figure 1. Common Medical Products Applications Using Polyurethane (Source: AI Generated)

WHY POLYURETHANE CONTINUES TO GAIN MARKET SHARE

Polyurethane offers a unique combination of properties that make it highly attractive for medical device applications.

High Mechanical Durability

Polyurethane exhibits excellent resistance to abrasion, puncture, flex fatigue, and repeated deformation. These characteristics are particularly valuable in devices that experience cyclic loading, inflation, pumping, or repeated patient interaction.

Biocompatibility and Hemocompatibility

Medical-grade polyurethane formulations demonstrate strong biocompatibility characteristics and are widely used in applications involving blood contact, tissue contact, and fluid handling. Common applications include catheters, wound care systems, fluid reservoirs, balloons, cuffs, and drug delivery devices.

Design Flexibility

Polyurethane can be formulated across a broad range of durometers, allowing designers to tailor mechanical properties to specific performance requirements. Material properties such as flexibility, transparency, barrier characteristics, and chemical resistance can be optimized without fundamentally changing the manufacturing approach.

Support for Emerging Medical Technologies

As wearable devices, minimally invasive therapies, and home healthcare technologies continue to expand, manufacturers increasingly require materials capable of supporting lightweight, flexible, and durable designs. Polyurethane aligns well with these evolving requirements.

UNDERSTANDING POLYURETHANE WELDABILITY

Successful welding begins with understanding how polyurethane behaves during thermal processing.

Unlike many semi-crystalline polymers, polyurethane typically exhibits more amorphous behavior, allowing it to soften and re-solidify in a predictable manner. This characteristic contributes to strong molecular fusion when appropriate welding parameters are applied.

Material Composition

Polyurethane formulations vary significantly among suppliers and product grades. Additives, lubricants, plasticizers, stabilizers, and processing aids can influence weldability, dielectric response, and long-term weld performance. Material selection should therefore consider not only functional performance but also compatibility with the intended joining process.

Moisture Sensitivity

Moisture control is particularly important for RF welding applications. Excess moisture can affect dielectric behavior and increase the risk of inconsistent heating, reduced weld quality, or process instability.

Film Thickness and Geometry

Uniform thickness is essential for achieving consistent energy distribution throughout the weld area. Significant thickness transitions can create localized hot spots or under-heated regions that negatively impact seal integrity.

Thin polyurethane films generally respond especially well to RF welding due to efficient dielectric heating and rapid energy transfer.

Thermal Characteristics

Polyurethane softens gradually over a range of temperatures rather than exhibiting an abrupt transition. This behavior provides a wider processing window and allows greater control of the fusion process.

However, excessive heat input can still cause degradation, discoloration, dimensional instability, or reduced mechanical performance.




Figure 2. Cross-Sectional Representation of a Polyurethane Overlap Weld Before and After RF Welding

WHY RF WELDING WORKS SO WELL WITH POLYURETHANE

RF welding is widely recognized as one of the most effective methods for joining polyurethane films and components.

Unlike conventional thermal welding processes that apply heat from the material surface inward, RF welding generates heat internally through dielectric excitation. As an alternating electromagnetic field passes through the material, molecular motion creates heat within the weld zone itself.

This approach offers several advantages.

Uniform Heating

Because energy is generated throughout the material thickness, heating is more uniform than many surface-heating technologies. This reduces the likelihood of scorching, surface damage, or incomplete fusion.

Strong Hermetic Seals

RF welding can produce highly reliable hermetic seals suitable for fluid containment applications such as reservoirs, bladders, bags, and inflatable components.

Repeatable Manufacturing

When process variables such as power, pressure, dwell time, and cooling time are controlled properly, RF welding provides highly repeatable results suitable for validated medical manufacturing environments.

Elimination of Adhesives

Adhesive bonding introduces additional materials, validation requirements, cure times, and contamination risks. RF welding eliminates many of these concerns while simplifying assembly operations.




Figure 3. RF Welding Process Cross Section

RF WELDING PROCESS CONSIDERATIONS

Achieving consistent weld quality requires careful control of several process variables.

RF Power

Power levels must be matched to the dielectric properties of the specific polyurethane formulation. Excessive power can cause overheating, while insufficient power may result in incomplete fusion.

Pressure Control

Proper electrode pressure promotes intimate contact between materials and supports molecular fusion. Excessive pressure can thin the weld area, while insufficient pressure may reduce seal strength.

Weld and Cool Times

Both heating and cooling phases influence final weld performance. Adequate cooling under pressure is often necessary to stabilize the weld and minimize distortion.

Tooling Design

Electrode geometry plays a critical role in energy distribution. Proper tooling design helps eliminate hot spots, improve seal consistency, and accommodate complex part geometries.

Multi-material assemblies may require specialized tooling configurations to ensure balanced energy delivery across varying thicknesses and material properties.

RF WELDING COMPARED TO OTHER JOINING METHODS

Several technologies are available for joining polyurethane components, each with advantages and limitations.

Ultrasonic Welding

Ultrasonic welding uses high-frequency mechanical vibrations to generate localized frictional heating. It can be highly effective for rigid and semi-rigid polyurethane components and offers extremely short cycle times. However, it may be less suitable for large film-based assemblies requiring broad hermetic seals.

Thermal and Impulse Welding

Thermal welding processes are commonly used for film sealing applications and can provide reliable results when properly controlled.

Because heat is introduced externally, achieving uniform temperature distribution may be more challenging than with RF welding, particularly for complex geometries.

Adhesive Bonding

Adhesive systems remain useful in certain applications but introduce additional validation requirements, longer cycle times, and potential contamination concerns. Bond performance may also be influenced by aging, environmental exposure, and cure conditions. For many polyurethane film applications, RF welding provides the most efficient combination of performance, repeatability, and manufacturing scalability.

Table 1. Comparison of Advanced Welding & Bonding Technologies

Feature	RF (Radio Frequency) Welding	Ultrasonic Welding	Thermal (Heat) Welding	Adhesive Bonding
Heat Source	High-frequency electromagnetic field	High-frequency mechanical vibration	Direct heat (hot jaw, laser, hot air)	None (chemical cure)
Bond Speed	Very fast (seconds)	Fastest (milliseconds to seconds)	Moderate (seconds to minutes)	Slowest (requires cure time)
Joint Strength	Strongest (hermetic, molecular bond)	High (strong molecular bond)	High (fused material bond)	Variable (relies on chemical adhesion)
Best Material Match	Polar plastics (TPU, PVC, PU)	Rigid plastics (ABS, PC, Nylon)	All thermoplastics (TPU, PE, PP)	Dissimilar materials (TPU to metal/wood)
Tooling & Setup Cost	High (custom dies required)	High (custom horns/fixtures)	Low to Moderate	Low (minimal initial equipment)

Figure 4. Table 1. Comparison of Advanced Welding & Bonding Technologies

DESIGNING FOR SUCCESSFUL RF WELDS

The most successful welding programs begin during product development rather than after design release.

Several design principles consistently improve weld quality and manufacturability:
- Maintain uniform film thickness wherever possible.
- Avoid abrupt transitions in geometry near weld zones.
- Select material durometers with weldability in mind.
- Design adequate weld lands and sealing surfaces.
- Minimize unnecessary stress concentrations.
- Consider tooling access early in development.
- Incorporate verification testing requirements into the design process.

Early collaboration between product designers and manufacturing engineers often reduces development risk, shortens validation timelines, and improves production yields.

VALIDATION AND PROCESS DEVELOPMENT

Medical device manufacturers must demonstrate that welding processes consistently produce assemblies meeting defined performance requirements.

Common validation activities include:

Mechanical Testing
Pull testing, peel testing, burst testing, and fatigue testing are frequently used to evaluate weld strength and durability.

Leak Testing
Leak testing verifies seal integrity and is particularly important for fluid containment applications.

Design of Experiments (DOE)
Structured experimentation helps identify critical process parameters and establish robust operating windows.

Process Validation
IQ, OQ, and PQ activities establish documented evidence that the welding process consistently produces acceptable products under normal production conditions.

The most successful validation programs combine process understanding with statistically meaningful testing methods to ensure long-term manufacturing stability.




Figure 5. Typical RF Welding Process Development and Validation Workflow

CONCLUSION

Polyurethane continues to gain acceptance throughout the medical device industry because it combines excellent biocompatibility, flexibility, durability, and design versatility. As device complexity increases and performance requirements become more demanding, manufacturers increasingly rely on joining technologies capable of producing reliable and repeatable assemblies.

Among available joining methods, RF welding remains one of the most effective solutions for polyurethane materials. Its ability to generate uniform internal heating, create strong hermetic seals, and eliminate adhesives makes it particularly valuable for fluid management, drug delivery, wearable, and inflatable medical devices.

Successful implementation requires more than selecting the right welding equipment. Material selection, part geometry, tooling design, process development, and validation planning all contribute to long-term manufacturing success.

By engaging manufacturing partners early in development and designing specifically for weld-ability, medical device companies can reduce risk, accelerate commercialization, and achieve robust production performance throughout the product lifecycle.

ABOUT DAVID BOWLING



David Bowling is a Senior RF Manufacturing Engineer with more than 35 years of experience in manufacturing engineering, process development, operations leadership, and advanced welding technologies. Throughout his career, he has held leadership positions spanning engineering, manufacturing, and operations management across medical device, industrial, and precision manufacturing environments.

David's expertise includes RF welding process development, manufacturing optimization, equipment validation, product launch support, and troubleshooting complex production challenges. He has led numerous initiatives involving advanced welding systems, automation, and process improvement programs designed to enhance product quality, operational efficiency, and manufacturing reliability.

ABOUT COMMAND MEDICAL PRODUCTS

COMMAND
MEDICAL PRODUCTS

Command Medical Products is a contract manufacturer of single use medical devices and custom medical kits, specializing in fluid management applications for infusion therapy, general surgery and laboratory and diagnostic markets.

Across its FDA registered, ISO 13485 certified facilities in Florida and Central America, Command provides precision tubing extrusion, advanced welding, ISO Class 7 clean room assembly, packaging and validation support.

Drawing on decades of medical manufacturing experience, Command combines disciplined quality systems and lean manufacturing with a transparent, highly collaborative approach, giving customers clear visibility, responsive communication and dependable execution throughout product development, manufacturing transfer, commercialization and ongoing production.

For more information visit commandmedical.com.